Message

Lisa Rector [Irector@nescaum.org] From:

3/25/2021 7:04:07 PM Sent:

To: Johnson, Steffan [johnson.steffan@epa.gov]

CC: Kelli O'Brien (kelli@clearstak.com) [kelli@clearstak.com]

Subject:

Attachments: Screen Shot 2021-03-25 at 1.52.38 PM.png

Okay, when you have a minute. Filter question for ASTM 3053 testing. The time between the end of the high fire and the start of the low fire was 18 minutes - very quick turnaround for multiple filter changes and leak checks. Even with multiple heat lamps on the filters before testing, the filter temp was below 80F for the first four minutes of the low fire test run - see screenshot. Any ideas on how this might be addressed?

Thanks!

Lisa

From: Johnson, Steffan < johnson.steffan@epa.gov>

Sent: Thursday, March 25, 2021 2:54 PM To: Lisa Rector < lrector@nescaum.org>

Subject: RE: call

Apologies, but I'm booked all afternoon.

From: Lisa Rector < irector@nescaum.org> Sent: Thursday, March 25, 2021 2:21 PM

To: Johnson, Steffan < johnson.steffan@epa.gov>

Subject: call

Any chance you can join our call now to discuss today's ASTM test? 855-738-3998.



Lisa Rector, Policy and Program Director at NESCAUM

89 South Street, Suite 602, Boston, Massachusetts, 02111 | 802.899.5306 | 617.259.2095 | Fax: 617.742.9162

| Irector@nescaum.org